

Quality-Assurance



Pagina	ID	Datum
1 van 1	SSP-230003	02 march 2012

We, from
Steril Safety Products

declare that our products meet the required standards, see specification below

Containerfilter with or without indicator DIN EN ISO 868-2
Article number SSP-500001, SSP-500002, SSP-500003, SSP-500004, SSP-500005.

Containerlabels and etikets, single or on rolls DIN EN ISO 11140-1
Article number SSP-300001, SSP-300002, SSP-300003, SSP-300004, SSP-300005,
SSP-300006, SSP-300010, SSP-300020, SSP-300021, SSP-300051, SSP-300071,
SSP-320001, SSP-320002, SSP-320003, SSP-320004,

Containerplomben DIN EN ISO 11140-1
Article number SSP-400001, SSP-400002, SSP-400003, SSP-400004, SSP-400005,
SSP-400006, SSP-400007

Batch cards / patient cards DIN EN ISO 11140-1
Article number SSP-600001, SSP-600002, SSP-600003, SSP-600004, SSP-600005.

Indicatorstrips DIN EN ISO 11140-1
Article number SSP-350001, SSP-350002, SSP-350003, SSP-350004

Bowie & Dick testpacks DIN EN ISO 11140-4
Article number SSP-800001, SSP-800002, SSP-800003

Bowie & Dick sheets DIN EN ISO 11140-3
Article number SSP-850001, SSP-850002

All products are being used as consumables for sterilisers.

The containerfilters are made of crepepaper and they fulfil the sterile Barriere regulations.

The products are compatible with containers of Aesculap, Wagner, Martin and Medicon.

The indicators contain no heavy metals.


Head of quality and ISO certifying
SSP produces sterilisation consumables